

Supplement

Title: 312, 316, 318 Users Supplement Issue: 1
Part Number: 1989445 Issue Date: 8/13
Print Date: July 2002 Page Count: 1
Revision/Date: 2, 2/06

This supplement contains information necessary to ensure the accuracy of the above Manual.

Change #1, 66328, 66329

On page 2, under **Environmental Conditions**, remove **Pollution Degree 2, Measurement Categories** and **Electromagnetic Compatibility**. Add the following:


Safety IEC 61010-1: 600 V CAT III / 1000 V CAT II, Pollution Degree 2

Electromagnetic Environment IEC 61326-1: Portable

Electromagnetic Compatibility Applies to use in Korea only. Class A Equipment (Industrial Broadcasting & Communication Equipment) ^[1]

[1] This product meets requirements for industrial (Class A) electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and is not to be used in homes.

On page 3, in the **Explanation of Symbols** table, replace the CAT II and CAT III with the following and add:

	Conforms to relevant South Korean EMC Standards.
CAT II	Measurement Category II is applicable to test and measuring circuits connected directly to utilization points (socket outlets and similar points) of the low-voltage MAINS installation.
CAT III	Measurement Category III is applicable to test and measuring circuits connected to the distribution part of the building's low-voltage MAINS installation.

On page 4, Under **General Specifications**, remove **Approval's** section.